



QUALITY CONTROL REPORT SURROGATE REPORT

Reference Number: 20-40299

Report Date: 11/25/20

Lab No	Analyte	Result	Qualifier	Units	Method	Limit
8270D_S201116 76453	2,4,6 - TRIBROMOPHENOL (Surr)	86		%	8270D	
	2 - FLUOROBIPHENYL (Surr)	79		%		
	2 - FLUOROPHENOL (Surr)	16		%		
	d5-PHENOL (Surr)	30		%		
	p-TERPHENYL-d14 (Surr)	91		%		
	d5-NITROBENZENE (Surr)	57		%		

***Notation:**

A surrogate is a pure compound added to a sample in the laboratory just before processing so that the overall efficiency of a meA surrogate is a pure compound added to a sample in the la

The Acceptance Limits (or Control Limits) approximate a 99% confidence interval around the mean recovery.